

**National Accreditation  
Agency of Ukraine**

**Approved by**  
the NAAU Decree  
№165-Я dated 14.02.2013

## **MANAGEMENT SYSTEM**

### *General document*

### ***“Peculiarities of application of the requirements of ДСТУ/IEC 17025 during accreditation of calibration laboratories”***

*(Approved and recommended for use by the  
Technical Accreditation Committee  
Minutes of the meeting dated December 21, 2013 No.33)*

<b>General document</b>			<b>NAAU</b>
<b>“Peculiarities of application of the requirements of ДСТУ/IEC 17025 during accreditation of calibration laboratories”</b>			
Revision 01 dated 14.02.2013	Developed by: Olha Khromenko	Verified by: Viktor Gorytskyi	Page 1
			Total number of pages 3

## **Peculiarities of the application of the requirements of ДСТУ ISO/IEC 17025 during accreditation of calibration laboratories**

1. At the time of assessment by NAAU for the purpose of accreditation against ДСТУ ISO/IEC 17025:2006 a calibration laboratory shall have an implemented management system in compliance with this Standard. A laboratory shall documentary complete its policies, systems, programs, procedures, methodologies and instructions in the amount, necessary for ensuring the quality of calibration results and their permanent use in work.

2. The evidences of competencies of personnel who carries out calibration of measurement techniques, assesses the results and completes the calibration certificates, shall be professional education and work experience in the field of metrology, documents confirming its authority and competencies, practical demonstration of specialist's gained skills and expertise during assessment.

3. A laboratory shall have a reference standards base, supplementary measurement techniques and equipment necessary for calibration in the relevant scope of accreditation.

4. A laboratory shall have and apply the procedure of measurement uncertainties' assessment during calibration for all types of measurement techniques in compliance with its scope of accreditation. Calibration certificates being issued by these laboratories shall contain the results of conducted calibration indicating the uncertainties.

5. A laboratory shall provide the evidences of traceability of its own standards and measurement devices against state standards of Ukraine or national standards of metrology institutes of other countries. The calibration certificates of laboratories' working standards shall include the information on uncertainties of measurement and traceability against relevant standards.

6. Laboratories shall have endorsed procedures and calibration methodologies, including calibration programs of their own working standards, supplementary measurement techniques and testing equipment.

7. Laboratories shall have a documentary evidence of assessment of suitability of calibration methodologies. Data on assessment of methodologies' suitability shall be updated in accordance with the envisaged intervals.

Head of Division on  
Laboratories' Accreditation

Olha Khromenko

<b>General document</b>			<b>NAAU</b>	
<b>“Peculiarities of application of the requirements of ДСТУ/IEC 17025 during accreditation of calibration laboratories”</b>				
Revision 01 dated 14.02.2013	Developed by: Olha Khromenko	Verified by: Viktor Gorytskyi	Page	2
			Total number of pages	2